

Section2	3.50	3.66	3.78	3.97
Section3	2.90	3.36	3.58	3.49
Section4	3.15	3.65	3.36	3.77
Section5	3.25	3.35	3.72	3.46
Section6	3.85	3.88	4.09	4.36
Section7	3.20	3.56	3.45	4.06

There may be some reasons for these errors' generating:

- (1) Every turn nonrecurring random rotary error has a great influence.
- (2) The transmission bands of the filter of TALYROND262 is 1 ~ 15 μ m and it is 1 ~ 50 μ m in our experiments. Maybe it is the reason for the difference between the real results and the measuring ones.
- (3) There is remnants radial rotary error in the results of error separation that is caused by the radial motion error of the shafting. The generating of the error in the experiment is caused by the above remnants error. Theoretically, the value of eccentricity of the separated roundness error function is zero. Actually, the value of eccentricity is numerously more than 0.02 μ m. The larger one can reach 0.1 ~ 0.25 μ m.
- (4) The measuring device has the adjustment error.

As a whole, the measuring results of roundness error tally well with the calibrated value.

IV. CONCLUSIONS

Simulation results show that when doing the micro-roughness fitting data, single-probe two-step error separation method take three times measurement to get the gyration error of fitted values is the same . That means when the measurement system's Gyration accuracy is very high, it's separation effect is very satisfying. It also can be concluded that the Single-probe two-step measurement method of roundness error separation technique can play a ideal separation effect on lathe spindle's bearing roller wear and thermal deformation etc system error.

When we add random errors to fitted value of three times measurement of gyration error, that means when rotary errors take different values, the outline of the roundness error is still kept consistent with the trends of the given error curve. Single-probe two-step method can not separate the random error off, but we can use the method of repeatedly measuring and taking average to decrease or basically eliminate random errors.

Several contrastive experiments of the on-line measurement are carried out in a grinding machine (MB1332A). The workpieces used in the experiments is measured by TALYROND 262 geometrical error measurement instrument. The results of these experiments successfully verify that the theory and mathematical models are correct. The contrast of the measuring results in these experiments proves that the Single-probe two-step method can realize the complete separation of all kinds of errors basically and this method is applicable to on-line measurement of a workpiece and it can also make the measurement accuracy very high.

Acknowledgement. This project was supported by National Science and Technology Major Projects of High-grade NC Machine Tools and Fundamental Manufacturing Equipment, P.R.China ((the project serial number is 2009ZX04001-062).

REFERENCES

- [1] Yasuo Aoki and Ozono Shigeo, "Research of Three-point Method Roundness Error measurement", *Precise machinery*, Vol. 32, No. 12, pp. 27-32. December 1966
- [2] R.R. Donaldson, "A simple method for separation spindle error from test ball roundness", *CIRP Annals*, Vol.21, No. 2, pp. 125–126. July 1972
- [3] D.J. Whitehouse, "Some theoretical aspects of error separation techniques in surface metrology", *J. Physics. E: Scien. Instru.*, Vol. 9, No. 7, pp. 531-536. July 1976
- [4] G. X. Zhang and R. K. Wang, "Four-point Method of Roundness and Spindle error Measurement", *CIRP Annals*, Vol. 42, No. 1, pp. 593-596, March 1993.
- [5] X. S. Zhu, B. Y. Xue, J. Y. Bei, "Interrelation and Application between Discrete Fourier Transform and Plural-real Fourier Transform", *Journal of Shanghai Jiaotong University*, Vol. 22, No. 4, pp. 1-12, April 1988.
- [6] E. W. Tyler, C. J. Evans and L.Z. Shao, "Uncertainly estimation for multiposition form error metrology", *Prec. Eng.*, Vol. 21-2, No.3, pp. 72-82, March 1997
- [7] W. Gao, S. Kiyono, and E. Satoh, "Precision measurement of multi-degree-of-freedom spindle errors using two- dimensional slope sensors", *Annals of CIRP*, Vol. 52, No.2, pp. 447-452, July 2002.
- [8] T. Ruedi and S. A. Juerg, "Primary roundness measuring machine", *Recent Developments in Traceable Dimensional Measurements III*, Vol. 2, pp. 1315-1321, 2005

- [9] Maisheng Hong, Zhonghuang Deng, Jianqiang Chen, et al. "Accurate time domain three point method for error separation of roundness", *Journal of Shanghai Jiaotong University*, Vol. 34, No.10, pp. 1317-1319 , October 2000.
- [10] Qi Wu, Hongbin Liu and Dejin Hu, "On-site measurement of roundness error based on error separation technique", *China Mechanical Engineering*, Vol. 8, No.17, pp.120-123, August 2006
- [11] Xunsheng Zhu, Guoliang Liu, Guoshong Li, et al. "A study of the theory for cylindricity and accuracy of rotational motion and straight motion error", *Journal of Shanghai Jiaotong University*, Vol. 26, No.4, pp.8-13, April 1992
- [12] Xunsheng Zhu, Dehui Pan, Bingyuan Xue, et al, "Two necessary conditions of the validity of the measured data for three-point EST and their applications", *Chinese Science Bulletin*, No.13, pp.79-80, July1986
- [13] Xunsheng Zhu, Bingyuan Xue, Jiyao Bei, "Definition of error motion of rotational spindle and separation of mandrel eccentricity and tilt", *Chinese Science Bulletin*, No.23, pp.1830-1833, November 1989
- [14] Yinbiao Guo, Jiang Guo, Xiao-long Ke, Yong-lu Zhu, "System Error Separation and Revision for Aspheric Lens On-Machine Measurement", *Nanotechnology and Precision Engineering*, Vol.8, No.5, pp.686-693, May 2010
- [15] Liang Yang , Xue-dong Jing , Kai He, "Embedded Roundness and Cylindricity Error Measuring Instruments", *Instrument Technique and Sensor*, No.3, pp.101-103, March 2011.
- [16] W. Gao, S. Kiyono and E. Satoh, "Precision measurement of multi-degree-of -freedom spindle errors using two-dimensional slope sensors", *Annals of CIRP*, Vol. 52, No. 2, pp. 447-452, March 2002.
- [17] O. Kamigaki, Y. Yamakawa, "Roundness measurement and its uncertainty in an international comparison", *Proc. of IMEKO 2000*, Oxford: Elsevier Science Ltd Oxford, Vol. 139- 144, 2000.
- [18] Lei Zhang, Ying Zhao and Yu Zhang, "Theory and Experiment of Three-probe Error Separation Technique", *Journal of Mechanical Engineering*, Vol. 45, No.6, pp. 256-261, June 2009.
- [19] L. Zhang, Y. Zhang and L. Ba, "Conditions of Precision Measurement for Error Separation by Multi-probe Method", *Advanced Materials Research*, Vol. 24-25, pp. 315-320, September 2007.
- [20] L. Zhang, X. M. Yang, L. Ba and Y. Zhang, "Parameter Selection Rules for Error Separation by Multi-probe Method", *Advanced Materials Research*. Vol. 53-54, pp. 421-427, October 2008.
- [21] Maisheng Hong, Ping Cain, "Analysis and Comparison of Multi-Step Error Separation Technique", *Journal of Shanghai Jiaotong University*, Vol. 38, No. 6, pp. 877-881, June 2004.
- [22] Lei Zhang, Xiumin Yang, Ming Hu, Lin Ba, and Dongrong Zheng, "Theoretical Studies and Simulation Analysis of the Time Domain Non-Symmetrical ConFigureuration Two-Probe

Method for Roundness Error Separation”, *Advanced Science Letters*, Vol.4, No.4-5, pp.1729–1734, September 2011

[23] Bangcheng Zhang, Xiaohong Yang, Di Wu, et al. “The application of two points method in the measurement of crankshaft roundness error”, *Modern Manufacturing Engineering*, No.6, pp. 75-77, June 2005

[24] Yu Zhang and Wenmin Li, “Measurement of roundness error by least squares mean circle method”, *Machinery Industry Standardization*, No.5, pp.3-5, May 1982